



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

Calimetro Perú SAC

***Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108***

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Thermodynamic, Chemical, Mechanical and Dimensional Calibration (As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

June 06, 2024

May 25, 2026

September 30, 2028

Accreditation No.:

Certificate No.:

126224

L26-471

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Thermodynamic	Thermo-Hygrometer (Humidity)	10 % RH to 90 % RH	1.8 % RH	Thermos-Hygrometer Operating Environment: Humidity And Temperature Chamber	PC-026 INACAL	F1, F3	F
Thermodynamic	Thermo-Hygrometer (Temperature)	10 °C to 40 °C	$(0.26 + 4.7 \times 10^{-3}T) ^\circ\text{C}$	Thermos-Hygrometer Operating Environment: Humidity And Temperature Chamber	PC-026 INACAL	F1, F3	F
Thermodynamic	Digital Thermometers	0 °C to 50 °C	$(0.03 + 5 \times 10^{-4}T) ^\circ\text{C}$	RTD Thermometer	PC-017 INACAL	F1, F3	F
Chemical	Gas Detector Carbon Monoxide (CO)	0.5×10^{-6} mol/mol to 45×10^{-6} mol/mol	$(0.017 \times 10^{-6} + 0.01\text{CO})$ mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection Dynamic Dilution System Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Carbon Monoxide (CO)	45×10^{-6} mol/mol to $1\,006 \times 10^{-6}$ mol/mol	$(0.226 \times 10^{-6} + 0.007\text{CO})$ mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Carbon Monoxide (CO)	100×10^{-6} mol/mol to $20\,100 \times 10^{-6}$ mol/mol	$(0.091 \times 10^{-6} + 0.02\text{CO})$ mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Gas Detector Nitric Oxide (NO)	10 x 10 ⁻⁹ mol/mol to 1 000 x 10 ⁻⁹ mol/mol	(0.12 x 10 ⁻⁹ + 0.012NO) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitric Oxide (NO)	1 x 10 ⁻⁶ mol/mol to 50 x 10 ⁻⁶ mol/mol	(0.016 x 10 ⁻⁶ + 0.01NO) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitric Oxide (NO)	50 x 10 ⁻⁶ mol/mol to 500 x 10 ⁻⁶ mol/mol	(0.207 x 10 ⁻⁶ + 0.008NO) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitric Oxide (NO)	500 x 10 ⁻⁶ mol/mol to 3 410 x 10 ⁻⁶ mol/mol	(1.23 x 10 ⁻⁶ + 0.01NO) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Sulfur Dioxide (SO ₂)	10 x 10 ⁻⁹ mol/mol to 1 000 x 10 ⁻⁹ mol/mol	(0.106 x 10 ⁻⁹ + 0.015SO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Sulfur Dioxide (SO ₂)	1 x 10 ⁻⁶ mol/mol to 50 x 10 ⁻⁶ mol/mol	(0.016 x 10 ⁻⁶ + 0.013SO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Gas Detector Sulfur Dioxide (SO ₂)	50 x 10 ⁻⁶ mol/mol to 505 x 10 ⁻⁶ mol/mol	(0.14 x 10 ⁻⁶ + 0.012SO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Sulfur Dioxide (SO ₂)	505 x 10 ⁻⁶ mol/mol to 3 411 x 10 ⁻⁶ mol/mol	(1.77 x 10 ⁻⁶ + 0.007 6SO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Methane (CH ₄)	25 x 10 ⁻⁶ mol/mol to 10 000 x10 ⁻⁶ mol/mol	(3.01 x 10 ⁻⁶ + 0.007CH ₄) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Methane (CH ₄)	12 500 x 10 ⁻⁶ mol/mol to 25 400 x 10 ⁻⁶ mol/mol	(0.03C x 10 ⁻⁶ + 0.006 x 10 ⁻⁶ CH ₄) mol/mol	EPA/NIST Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Hydrogen Sulfide (H ₂ S)	10 x 10 ⁻⁹ mol/mol to 1 000 x 10 ⁻⁹ mol/mol	(0.087x 10 ⁻⁹ + 0.021H ₂ S) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Hydrogen Sulfide (H ₂ S)	1 x 10 ⁻⁶ mol/mol to 200 x 10 ⁻⁶ mol/mol	(0.007 x 10 ⁻⁶ + 0.021H ₂ S) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Gas Detector Hydrogen Sulfide (H ₂ S)	200 8 x 10 ⁻⁶ mol/mol	(0.02 x 10 ⁻⁶ H ₂ S) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Oxygen (O ₂)	0.1 x 10 ⁻² mol/mol to 21 x 10 ⁻² mol/mol	(0.04 x 10 ⁻² + 0.004O ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Carbon Dioxide (CO ₂)	0.02 x 10 ⁻² mol/mol to 14 x 10 ⁻² mol/mol	(0.023 x 10 ⁻² + 0.007CO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitrogen Dioxide (NO ₂)	1.5 x 10 ⁻⁶ mol/mol to 450 x 10 ⁻⁶ mol/mol	(0.02 x 10 ⁻⁶ + 0.012NO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's P20 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitrogen Dioxide (NO ₂)	300 3 x 10 ⁻⁶ mol/mol	(30.036 x 10 ⁻⁶ NO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's P20 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Nitrogen Dioxide (NO ₂)	50 x 10 ⁻⁹ mol/mol to 500 x 10 ⁻⁹ mol/mol	(2.23 x 10 ⁻⁹ + 0.034NO ₂) mol/mol	EPA Certified Gas Dynamic Dilution System (MFC) Direct Injection	P19 Internal SOP's P20 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Ozone (O ₃)	50 x 10 ⁻⁹ mol/mol to 500 x 10 ⁻⁹ mol/mol	(1.91 x 10 ⁻⁹ + 0.032O ₃) mol/mol	Photometer (Ozone Generator)	P21 Internal SOP's	F1, F3	F, O



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Gas Detector Isobutylene (C ₄ H ₈)	10 x 10 ⁻⁶ mol/mol	(0.06 x 10 ⁻⁶ C ₄ H ₈) mol/mol	EPA/NIST Certified Gas Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	Gas Detector Isobutylene (C ₄ H ₈)	100 x 10 ⁻⁶ mol/mol	(0.03 x 10 ⁻⁶ C ₄ H ₈) mol/mol	EPA/NIST Certified Gas Direct Injection	P19 Internal SOP's	F1, F3	F, O
Chemical	pH Meters	4 pH	0.012 pH	Buffer MRC	PC-020 INACAL	F1, F3	F, O
Chemical	pH Meters	7 pH	0.012 pH	Buffer MRC	PC-020 INACAL	F1, F3	F, O
Chemical	pH Meters	10 pH	0.012 pH	Buffer MRC	PC-020 INACAL	F1, F3	F, O
Chemical	ORP Meter (Redox Potential Meter)	229 mV	1.8 mV	Solution MRC	P34 Internal SOP's	F1, F3	F, O
Chemical	ORP Meter (Redox Potential meter)	470 mV	4.5 mV	Solution MRC	P34 Internal SOP's	F1, F3	F, O
Chemical	Conductivity Meter	10 μ S/cm	0.6 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Conductivity Meter	100 μ S/cm	2.2 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Conductivity Meter	1 413 μ S/cm	4.8 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Conductivity Meter	10 000 μ S/cm	50 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Conductivity Meter	12 880 μ S/cm	50 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Conductivity Meter	100 000 μ S/cm	370 μ S/cm	Solution MRC	PC-022 INACAL	F1, F3	F, O
Chemical	Oximeters (Dissolved Oxygen)	Up to 9.4 mg/L	0.06 mg/L	Thermometer Barometer Solution MRC	P28 Internal SOP's	F1, F3	F



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Turbidimeter (Turbidity)	10 NTU	0.1 NTU	Solution MRC	P29 Internal SOP's	F1, F3	F, O
Chemical	Turbidimeter (Turbidity)	20 NTU	0.2 NTU	Solution MRC	P29 Internal SOP's	F1, F3	F, O
Chemical	Turbidimeter (Turbidity)	100 NTU	2 NTU	Solution MRC	P29 Internal SOP's	F1, F3	F, O
Chemical	Turbidimeter (Turbidity)	400 NTU	4 NTU	Solution MRC	P29 Internal SOP's	F1, F3	F, O
Chemical	Turbidimeter (Turbidity)	800 NTU	8 NTU	Solution MRC	P29 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	0.2 mg/L	0.04 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	0.5 mg/L	0.04 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	0.9 mg/L	0.04 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	1.5 mg/L	0.04 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	2 mg/L	0.04 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	3 mg/L	0.08 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	4 mg/L	0.08 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Chemical	Chlorine Analyzer	5 mg/L	0.08 mg/L	Solution MRC	P30 Internal SOP's	F1, F3	F, O
Mechanical	Anemometer	0.5 m/s to 25 m/s	(0.2 + 0.028 4Va) m/s	Digital Anemometer wind Tunnel	P15 Internal SOP's	F1, F3	F
Mechanical	Atmospheric Barometer	550 hPa to 1 100 hPa	0.50 hPa	Digital Barometer Pressure Chamber	PC-024 INACAL	F1, F3	F



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Ambient Pressure Sensors	413 mmHg to 788 mmHg	0.41 mmHg	Digital Barometer Pressure Generating Pump	P35 Internal SOP's	F1, F3	F, O
Mechanical	Liquid Column Manometer	0 inH ₂ O to 160 inH ₂ O	0.06 inH ₂ O	Digital Manometer Pressure Generating Pump	ME-021	F1, F3	F, O
Mechanical	Liquid Column Manometer	0 kPa to 40 kPa	0.015 kPa	Digital Manometer Pressure Generating Pump	ME-021	F1, F3	F, O
Mechanical	Differential Pressure Gauges	0 inH ₂ O to 160 inH ₂ O	0.03 inH ₂ O	Digital Manometer Pressure Generating Pump	ME-020	F1, F3	F, O
Mechanical	Differential Pressure Gauges	0 kPa to 40 kPa	0.008 kPa	Digital Manometer Pressure Generating Pump	ME-020	F1, F3	F, O
Mechanical	Flow Meters, Flowmeters, Particle Samplers, Rotameters, Flowmeters	15 cm ³ /min to 500 cm ³ /min	(0.1 + 0.002 2Q) cm ³ /min	Flow Meter Variflow Calibrator	ME-009 - CEM P31 Internal SOP's	F1, F3	F, O
Mechanical	Flow Meters, Flowmeters, Particle Samplers, Rotameters, Flowmeters	0.5 L/min to 30 L/min	(0.002 + 0.003 2Q) L/min	Flow Meter Variflow Calibrator	ME-009 - CEM P31 Internal SOP's	F1, F3	F, O
Mechanical	Flow Meters (High Volume Particulate Matter Samplers)	0.9 m ³ /min to 1.9 m ³ /min	0.033 m ³ /min	Flow Meter Variflow Calibrator	ME-009 - CEM P31 Internal SOP's	F1, F3	F, O



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108
Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Liquid Flow Meter (Rain Gauges)	Up to 480 mm	0.13 mm	Graduated Container – Rain Gauge Calibrator	P32 Internal SOP's	F1, F3	F, O
Dimensional	Wind Vane & Wind Direction Sensor	Up to 360°	0.06°	Goniometer	P33 Internal SOP's	F1, F3	F

- The CMC (Calibration and Measurement Capability) is expressed in terms of measurement instrument/aspect being calibrated, range, expanded measurement uncertainty, equipment, and method/procedure. The expanded measurement uncertainty stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the measurement uncertainty included on this scope for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
- The laboratory's range of calibration capability for all disciplines for which it is accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
- Location of activity:

Location Code	Location
F	Conformity assessment activity is performed at the CAB's fixed facility
O	Conformity assessment activity is performed onsite at the CAB's customer location
- Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratory's fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratory's fixed location.



Certificate of Accreditation: Supplement

Calimetro Perú SAC

Av Universitaria N° 1593 Mz C Lote 4, Daniel Alcides Carrión
San Martín de Porres, Lima, Perú, C.P. 15108

Contact Name: Isaias Curi. Phone: (51)966-305-560

Accreditation is granted to the facility to perform the following conformity assessment activities:

5. The term T represents temperature in °C or °F as appropriate to the uncertainty statement.
6. The term Va represents the air velocity in meters/second or feet/second (including SI multiple and submultiple units) as appropriate to the uncertainty statement.
7. The term Q represents Flow Rate in L/min including SI multiple and submultiple units) as appropriate to the uncertainty statement.
8. Flex Codes

F0: When no flexibility is identified. There are no changes to items calibrated, characteristics identified or versions of methods except for updating to the most recent version of a standard method after verification.

F1: The laboratory has the capability to introduce a new instrument, quantity, or gauge for an accredited calibration method.

F2: The laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope

F3: The laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope

F4: The laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using the same Calibration Equipment or Reference Standards identified on the scope for the same parameter, component, or analyte identified on the line item of the scope.